## Annex A: Specifications and Compliance Sheet

Vendors are required to fill in the 'For Vendors to complete' section

BDFA/Q/0019/2024

## A.1 Temperature Data Logger

|    |                                      |   | For Vendors to complete |                     |
|----|--------------------------------------|---|-------------------------|---------------------|
| No | Descriptions                         | Specifications  | Comply<br>(Yes/No)      | State Your Proposal |
| 1  | Parameter                            | Temperature - reading in Celsius, to 2 digital place.   |                         |                     |
| 2  | Туре                                 | Digital   |                         |                     |
| 3  | Display                              | LCD   |                         |                     |
| 4  | Usage type                           | Multiple use  |                         |                     |
| 5  | Connection<br>Interface              | At least USB 2.0 or above   |                         |                     |
| 6  | Measurement<br>Range                 | In the range of between -40°C to +70°C with plus and minus 15% variance   |                         |                     |
| 7  | Recording<br>Capability              | 16,000 readings   |                         |                     |
| 8  | Log Interval                         | 1 minute – 24 hours   |                         |                     |
| 9  | IP Rating                            | At least IP65 or above  |                         |                     |
| 10 | Operating<br>System<br>Compatibility | Microsoft Window 8/10   |                         |                     |
| 11 | Report Format                        | PDF / CSV report  |                         |                     |
| 12 | Battery                              | Preferable with replaceable battery   |                         |                     |
| 13 | Calibration                          | Each unit has to be calibrated and certified<br>with validity of at least 12 months. Each unit<br>must be accompanied with certificate of<br>calibration. |                         |                     |
| 14 | Quantity<br>required                 | 15  |                         |                     |
| 15 | Delivery Date                        | 30/01/2025  |                         |                     |

Note: The Specifications and Compliance Sheet must be signed and stamped